Certificate
of the following UK Regulation

Registered No.:
ND1105220005

S.I. 618 The Medical Device Regulations 2002
Class I Device

Reference of applicant	Date of application	File reference	Test report No.	Date of issue	Expiry date
-	12.09.2022	TWR2209011	TWR2209011 001	19.10.2022	18.10.2027

This is to certify that the following product complies to the related provisions of the above mentioned UK Regulation and the following standards:

Product: Linear actuator

Type designation: LAS series (details model numbers see Appendix 1)

Serial No.: --

Applicant: HIWIN MIKROSYSTEM CORP.
No. 6, Jingke Central Rd., Taichung Precision Machinery Park, Taichung City 40852, Taiwan

Standard(s): BS EN 60601-1:2006, BS EN 60601-1:2006/A1:2013, BS EN 60601-1:2006/A2:2021

This Certificate is based on the voluntary evaluation of sample(s) of the above mentioned product. It does not imply an assessment of the production and it does not permit the use of a mark of conformity or of a safety mark of the TÜV NORD Group. This is to certify that the tested sample is in compliance with the relevant requirements referred in Part II of The Medical Device Regulations 2002. This certificate can be used by the certificate holder in connection with the UKCA declaration of conformity indicating conformity according to The Medical Device Regulations 2002.

Certification Department

Email:info.pc@tuv-nord.com
www.tuv-nord.com

The certificate can be verified by Online certification search in TUV NORD Taiwan Co., Ltd. website: www.tuvnord.com.tw.
TAF ISO 17065 accredited no. PC094 for certification. This certificate and all appendices must be reproduced in full.
The UKCA marking may be affixed on the product if all relevant and effective Regulations are complied with.
The TÜV NORD Group is active in more than 70 countries globally, which including TÜV NORD CERT GmbH (NB 0044), TÜV NORD SYSTEMS GmbH & CO. KG(NB 0045), TÜV NORD BALTIK LTD (NB 1409),TÜV NORD Sweden AB (NB 2529), TÜV UK Ltd. (AB 0879)...etc.

Certificate
of the following UK Regulation


Registered No.:
ND1105220005

S.I. 618 The Medical Device Regulations 2002
Class I Device


Reference of applicant	Date of application	File reference	Test report No.	Date of issue	Expiry date
-	12.09.2022	TWR2209011	TWR2209011 001	19.10.2022	18.10.2027

Model list

Sub series	Model no.	Thrust max. (N)	Pulling max. (N)	Holding max. (N)	Speed (mm/s) Load = Max. / Load = 0	Motor type: rpm	Current max. (A)	
							DC 12V	DC 24V
LAS1	LAS1-1	1200	1200	800	8/12	AM6-0-012G: 7750 AM6-0-024G: 7000	6	2.5
	LAS1-2	600	600	300	16/25		6	3



LAS2	LAS2-1	1800	1200	1800	4.5/7	AM4-0-012G: 4000 AM4-0-024G: 4400	6	3
	LAS2-2	1200	1200	1000	8/15		6	4



Certification Department

Email:info.pc@tuv-nord.com
www.tuv-nord.com

The certificate can be verified by Online certification search in TUV NORD Taiwan Co., Ltd. website: www.tuvnord.com.tw.
TAF ISO 17065 accredited no. PC094 for certification. This certificate and all appendices must be reproduced in full.
The UKCA marking may be affixed on the product if all relevant and effective Regulations are complied with.
The TÜV NORD Group is active in more than 70 countries globally, which including TÜV NORD CERT GmbH (NB 0044), TÜV NORD SYSTEMS GmbH & CO. KG(NB 0045), TÜV NORD BALTIK LTD (NB 1409),TÜV NORD Sweden AB (NB 2529), TÜV UK Ltd. (AB 0879)...etc.

Certificate
of the following UK Regulation


Registered No.:
ND1105220005

S.I. 618 The Medical Device Regulations 2002
Class I Device

Reference of applicant	Date of application	File reference	Test report No.	Date of issue	Expiry date
-	12.09.2022	TWR2209011	TWR2209011 001	19.10.2022	18.10.2027

Model list

Sub series	Model no.	Thrust max. (N)	Pulling max. (N)	Holding max. (N)	Speed (mm/s) Load = Max. / Load = 0	Motor type: rpm	Current max. (A)	
							DC 12V	DC 24V
LAS3	LAS3-1	1200	1200	800	8/12	AM6-0-012G: 7750 AM6-0-024G: 7000	6	2.5
	LAS3-2	600	600	300	16/25		6	3



Certification Department

Email:info.pc@tuv-nord.com
www.tuv-nord.com

The certificate can be verified by Online certification search in TUV NORD Taiwan Co., Ltd. website: www.tuvnord.com.tw.
TAF ISO 17065 accredited no. PC094 for certification. This certificate and all appendices must be reproduced in full.
The UKCA marking may be affixed on the product if all relevant and effective Regulations are complied with.
The TÜV NORD Group is active in more than 70 countries globally, which including TÜV NORD CERT GmbH (NB 0044), TÜV NORD SYSTEMS GmbH & CO. KG(NB 0045), TÜV NORD BALTIK LTD (NB 1409),TÜV NORD Sweden AB (NB 2529), TÜV UK Ltd. (AB 0879)...etc.

Certificate
of the following UK Regulation


Registered No.:
ND1105220005

S.I. 618 The Medical Device Regulations 2002
Class I Device

Reference of applicant	Date of application	File reference	Test report No.	Date of issue	Expiry date
-	12.09.2022	TWR2209011	TWR2209011 001	19.10.2022	18.10.2027

Model list

Sub series	Model no.	Thrust max. (N)	Pulling max. (N)	Holding max. (N)	Speed (mm/s) Load = Max. / Load = 0	Motor type: rpm	Current max. (A)	
							DC 12V	DC 24V
LAS4	LAS4-1	800	800	600	10/15	AM6-3-012: 155 or AM6-3-024: 155	5	2.3
	LAS4-2	300	300	200	30/46	AM6-3-012: 300 or AM6-3-024: 310	6	3.6



Certification Department

Email:info.pc@tuv-nord.com
www.tuv-nord.com

The certificate can be verified by Online certification search in TUV NORD Taiwan Co., Ltd. website: www.tuvnord.com.tw.
TAF ISO 17065 accredited no. PC094 for certification. This certificate and all appendices must be reproduced in full.
The UKCA marking may be affixed on the product if all relevant and effective Regulations are complied with.
The TÜV NORD Group is active in more than 70 countries globally, which including TÜV NORD CERT GmbH (NB 0044), TÜV NORD SYSTEMS GmbH & CO. KG(NB 0045), TÜV NORD BALTIK LTD (NB 1409),TÜV NORD Sweden AB (NB 2529), TÜV UK Ltd. (AB 0879)...etc.

Certificate
of the following UK Regulation


Registered No.:
ND1105220005

S.I. 618 The Medical Device Regulations 2002
Class I Device

Reference of applicant	Date of application	File reference	Test report No.	Date of issue	Expiry date
-	12.09.2022	TWR2209011	TWR2209011 001	19.10.2022	18.10.2027

Model list

Sub series	Model no.	Thrust max. (N)	Pulling max. (N)	Holding max. (N)	Speed (mm/s) Load = Max. / Load = 0	Motor type: rpm	Current max. (A)	
							DC 12V	DC 24V
LAS5	LAS5-1	800	800	600	2.5/4	IG-28GM	--	1
	LAS5-2	300	300	200	6/7.5		--	1
	LAS5A-1	2000	2000	1500	2.5/4	IG-36PGM	--	2.3
	LAS5A-1	1200	1200	800	7/8.5		--	2.3



Certification Department

Email:info.pc@tuv-nord.com
www.tuv-nord.com

The certificate can be verified by Online certification search in TUV NORD Taiwan Co., Ltd. website: www.tuvnord.com.tw.
TAF ISO 17065 accredited no. PC094 for certification. This certificate and all appendices must be reproduced in full.
The UKCA marking may be affixed on the product if all relevant and effective Regulations are complied with.
The TÜV NORD Group is active in more than 70 countries globally, which including TÜV NORD CERT GmbH (NB 0044), TÜV NORD SYSTEMS GmbH & CO. KG(NB 0045), TÜV NORD BALTIK LTD (NB 1409),TÜV NORD Sweden AB (NB 2529), TÜV UK Ltd. (AB 0879)...etc.